



# Maintenance of HV Measurement Technology – Calibration and Insulation Test

CSM web seminars

**CSM** **Xplained**  
measurement technology




Innovative Measurement and Data Technology



## Calibration

Expiring in  
September 2022




## DAkkS

Requested by the  
customer



## HV Insulation Test

Expiring in  
September 2022




## Time

How can the tests still  
be carried out in time?

# Calibration

## Definition and procedure


- **Comparison** of a measured value (display value) with a reference value



# Calibration

## Definition and procedure

- ▶ Comparison of a measured value (display value) with a reference value
- ▶ Difference between nominal and actual value = **measurement deviation** of the calibration object






# Calibration

## Definition and procedure

- ▶ Comparison of a measured value (display value) with a reference value
- ▶ Difference between nominal and actual value = measurement deviation of the calibration object
- ▶ **To be performed separately for all measurement channels / measurement ranges**






# Calibration

## Definition and procedure

- ▶ Comparison of a measured value (display value) with a reference value
- ▶ Difference between nominal and actual value = measurement deviation of the calibration object
- ▶ To be performed separately for all measurement channels / measurement ranges
- ▶ Observe **measurement uncertainty** due to calibration setup



# Calibration

## Calibration certificate



CSM GmbH  
Computer-Systeme-Messtechnik

Kalibrierschein / Calibration certificate

erstellt durch das Kalibrierlaboratorium  
issued by the calibration laboratory

Kalibrierlabor CSM GmbH

Kalibrierung  
akkreditiert  
DIN EN ISO 9001

Mitgliedsnummer  
De 123456789

Gegenstand  
Herstellung  
Typ  
Seriennummer  
AI 1234

Messergebnisse  
Measurement results

Thermospannung Typ K  
Thermoelectric voltage type K

Calibration certificate – page 4

| Kanal<br>Channel | Referenzwert<br>Reference value | Messwert<br>Measured value | Messunsicherheit<br>Measurement uncertainty | Abweichung<br>Deviation | Toleranz<br>Tolerance | Abweichung / Toleranz<br>Deviation / Tolerance | Ergebnis<br>Result |
|------------------|---------------------------------|----------------------------|---------------------------------------------|-------------------------|-----------------------|------------------------------------------------|--------------------|
| 1                | -95,00 °C                       | -95,17 °C                  | 0,30 K                                      | -0,17 K                 | 0,65 K                | -26,2 %                                        | ✓                  |
| 1                | 0,00 °C                         | -0,13 °C                   | 0,30 K                                      | -0,13 K                 | 0,60 K                | -21,7 %                                        | ✓                  |
| 1                | 100,00 °C                       | 99,88 °C                   | 0,30 K                                      | -0,12 K                 | 0,65 K                | -18,5 %                                        | ✓                  |
| 1                | 1.000,00 °C                     | 999,95 °C                  | 0,30 K                                      | -0,05 K                 | 1,10 K                | -4,5 %                                         | ✓                  |
| 1                | 1.370,00 °C                     | 1.370,02 °C                | 0,30 K                                      | 0,02 K                  | 1,29 K                | 1,6 %                                          | ✓                  |
| 2                | -95,00 °C                       | -95,08 °C                  | 0,30 K                                      | -0,08 K                 | 0,65 K                | -12,3 %                                        | ✓                  |
| 2                | 0,00 °C                         | -0,09 °C                   | 0,30 K                                      | -0,09 K                 | 0,60 K                | -15,0 %                                        | ✓                  |
| 2                | 100,00 °C                       | 99,97 °C                   | 0,30 K                                      | -0,03 K                 | 0,65 K                | -4,6 %                                         | ✓                  |
| 2                | 1.000,00 °C                     | 999,97 °C                  | 0,30 K                                      | -0,03 K                 | 1,10 K                | -2,7 %                                         | ✓                  |
| 2                | 1.370,00 °C                     | 1.370,00 °C                | 0,30 K                                      | 0,00 K                  | 1,29 K                | 0,0 %                                          | ✓                  |
| 3                | -95,00 °C                       | -95,10 °C                  | 0,30 K                                      | -0,10 K                 | 0,65 K                | -15,4 %                                        | ✓                  |
| 3                | 0,00 °C                         | -0,10 °C                   | 0,30 K                                      | -0,10 K                 | 0,60 K                | -16,7 %                                        | ✓                  |
| 3                | 100,00 °C                       | 99,99 °C                   | 0,30 K                                      | -0,01 K                 | 0,65 K                | -1,5 %                                         | ✓                  |
| 3                | 1.000,00 °C                     | 1.000,01 °C                | 0,30 K                                      | 0,01 K                  | 1,10 K                | 0,9 %                                          | ✓                  |
| 3                | 1.370,00 °C                     | 1.370,05 °C                | 0,30 K                                      | 0,05 K                  | 1,29 K                | 3,9 %                                          | ✓                  |
| 4                | -95,00 °C                       | -95,16 °C                  | 0,30 K                                      | -0,16 K                 | 0,65 K                | -24,6 %                                        | ✓                  |
| 4                | 0,00 °C                         | -0,14 °C                   | 0,30 K                                      | -0,14 K                 | 0,60 K                | -23,3 %                                        | ✓                  |
| 4                | 100,00 °C                       | 99,94 °C                   | 0,30 K                                      | -0,06 K                 | 0,65 K                | -9,2 %                                         | ✓                  |
| 4                | 1.000,00 °C                     | 999,97 °C                  | 0,30 K                                      | -0,03 K                 | 1,10 K                | -2,7 %                                         | ✓                  |
| 4                | 1.370,00 °C                     | 1.370,02 °C                | 0,30 K                                      | 0,02 K                  | 1,29 K                | 1,6 %                                          | ✓                  |

- ▶ Reference value
- ▶ Measurement derivation
- ▶ Measurement channels / measurement ranges
- ▶ Measurement uncertainty



# Calibration

## Conformity Statement



CSM GmbH  
Computer-Systeme-Messtechnik

### Kalibrierschein / Calibration certificate

erstellt durch das Kalibrierlaboratorium  
issued by the calibration laboratory

Kalibrierlaboratorium CSM GmbH

### Kalibrierung / Calibration Messergebnisse Measurement results

### Thermospannung Typ K Thermoelectric voltage type K

| Kanal<br>Channel | Referenzwert<br>Reference value | Messwert<br>Measured value | Messunsicherheit<br>Measurement uncertainty | Abweichung<br>Deviation | Toleranz<br>Tolerance | Abweichung / Toleranz<br>Deviation / Tolerance | Ergebnis<br>Result |
|------------------|---------------------------------|----------------------------|---------------------------------------------|-------------------------|-----------------------|------------------------------------------------|--------------------|
| 1                | -95,00 °C                       | -95,17 °C                  | 0,30 K                                      | -0,17 K                 | 0,65 K                | -26,2 %                                        | ✓                  |
| 1                | 0,00 °C                         | -0,13 °C                   | 0,30 K                                      | -0,13 K                 | 0,60 K                | -21,7 %                                        | ✓                  |
| 1                | 100,00 °C                       | 99,88 °C                   | 0,30 K                                      | -0,12 K                 | 0,65 K                | -18,5 %                                        | ✓                  |
| 1                | 1.000,00 °C                     | 999,95 °C                  | 0,30 K                                      | -0,05 K                 | 1,10 K                | -4,5 %                                         | ✓                  |
| 1                | 1.370,00 °C                     | 1.370,02 °C                | 0,30 K                                      | 0,02 K                  | 1,29 K                | 1,6 %                                          | ✓                  |
| 2                | -95,00 °C                       | -95,08 °C                  | 0,30 K                                      | -0,08 K                 | 0,65 K                | -12,3 %                                        | ✓                  |
| 2                | 0,00 °C                         | -0,09 °C                   | 0,30 K                                      | -0,09 K                 | 0,60 K                | -15,0 %                                        | ✓                  |
| 2                | 100,00 °C                       | 99,97 °C                   | 0,30 K                                      | -0,03 K                 | 0,65 K                | -4,6 %                                         | ✓                  |
| 2                | 1.000,00 °C                     | 999,97 °C                  | 0,30 K                                      | -0,03 K                 | 1,10 K                | -2,7 %                                         | ✓                  |
| 2                | 1.370,00 °C                     | 1.370,00 °C                | 0,30 K                                      | 0,00 K                  | 1,29 K                | 0,0 %                                          | ✓                  |
| 3                | -95,00 °C                       | -95,10 °C                  | 0,30 K                                      | -0,10 K                 | 0,65 K                | -15,4 %                                        | ✓                  |
| 3                | 0,00 °C                         | -0,10 °C                   | 0,30 K                                      | -0,10 K                 | 0,60 K                | -16,7 %                                        | ✓                  |
| 3                | 100,00 °C                       | 99,99 °C                   | 0,30 K                                      | -0,01 K                 | 0,65 K                | -1,5 %                                         | ✓                  |
| 3                | 1.000,00 °C                     | 1.000,01 °C                | 0,30 K                                      | 0,01 K                  | 1,10 K                | 0,9 %                                          | ✓                  |
| 3                | 1.370,00 °C                     | 1.370,05 °C                | 0,30 K                                      | 0,05 K                  | 1,29 K                | 3,9 %                                          | ✓                  |
| 4                | -95,00 °C                       | -95,16 °C                  | 0,30 K                                      | -0,16 K                 | 0,65 K                | -24,6 %                                        | ✓                  |
| 4                | 0,00 °C                         | -0,14 °C                   | 0,30 K                                      | -0,14 K                 | 0,60 K                | -23,3 %                                        | ✓                  |
| 4                | 100,00 °C                       | 99,94 °C                   | 0,30 K                                      | -0,06 K                 | 0,65 K                | -9,2 %                                         | ✓                  |
| 4                | 1.000,00 °C                     | 999,97 °C                  | 0,30 K                                      | -0,03 K                 | 1,10 K                | -2,7 %                                         | ✓                  |
| 4                | 1.370,00 °C                     | 1.370,02 °C                | 0,30 K                                      | 0,02 K                  | 1,29 K                | 1,6 %                                          | ✓                  |

Calibration vertificate – page 4

- ▶ Reference value
- ▶ Measurement derivation
- ▶ Measurement channels / measurement ranges
- ▶ Measurement uncertainty
- ▶ **Conformity statement**

Comparison of the calibration values with limit values from specifications of the test specimen including the measurement uncertainty



# Calibration Conformity Statement



CSM GmbH  
Computer-Systeme-Messtechnik

## Kalibrierschein

erstellt durch das Kalibr  
issued by the calibration

## Kalibrierlabor

akkreditiert nach DIN E  
accredited according to

Mitglied im  
Member of

## Deutscher

Gegenstand  
Object

Hersteller  
Manufacturer

Typ  
Type

Seriennummer  
Serial number

Auftraggeber  
Customer

Auftragsnummer  
Order number

Anzahl der Seiten  
Number of pages

Datum der Kalibrierung  
Date of calibration

Dieser Kalibrierschein  
ausdrücklichen Gene  
Verantwortlichen in KI  
This calibration certifi-  
cates with the fi-

Datum und Freigabe  
Date and approval of the

CSM Computer-Systeme-Messtechnik  
Kraftfahrtstraße 36-7  
Es gelten ausschließlich die  
Bedingungen der Allgemeinen

## Konformitätsaussage

### Conformity statement

Die Messergebnisse der Kalibrierung werden auf Konformität mit der Herstellerspezifikation des Kalibriergegenstands unter Berücksichtigung der erweiterten Messunsicherheit ( $k = 2$ , mit Überdeckungswahrscheinlichkeit 95 %) in Anlehnung an DIN EN ISO 14253-1:2018-07 bewertet. Die Prüfung erfolgt mit Schutzbandfaktor  $g = 2$ , wodurch eine Konformitätswahrscheinlichkeit von 95 % und höher erreicht wird. Das Bewertungsergebnis wird für jeden einzelnen Messwert im Kalibrierschein wie folgt angezeigt:

The measurement results of the calibration are evaluated for conformity with the manufacturer's specification of the calibration object under consideration of the extended measurement uncertainty ( $k = 2$ , with coverage probability 95 %) referring to DIN EN ISO 14253-1:2018-07. The conformity assessment is carried out with guard band factor  $g = 2$ , which results in a conformance probability of 95 % and higher. The evaluation result for each individual measured value is given in the calibration certificate as follows:

| Bewertungsergebnis<br>Evaluation result                                                                                                        | Kennzeichnung<br>Marking | Exemplarisches Messergebnis<br>Measurement result example |
|------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------|-----------------------------------------------------------|
| Innerhalb der Spezifikation<br>Within specification                                                                                            | ✓                        |                                                           |
| Keine Angabe möglich, Messwert im Unsicherheitsbereich<br>No statement possible, measured value in the uncertainty range                       | —                        |                                                           |
| Messwert außerhalb des Toleranzbereichs und im Unsicherheitsbereich<br>Measured value outside the tolerance range and in the uncertainty range | ✗                        |                                                           |
| Außerhalb der Spezifikation<br>Out of specification                                                                                            | ✗                        |                                                           |

Toleranzgrenzen / Tolerance limits  
 Referenzwert / Reference value  
 Messwert / Measured value  
 Messunsicherheit / Measurement uncertainty

## Zusammenfassung je Kanal über alle Messwerte

### Summary by channel over all measured values

| Messgröße<br>Measurand | Thermospannung Typ K<br>Thermoelectric voltage type K |   |   |   |
|------------------------|-------------------------------------------------------|---|---|---|
| Kanal<br>Channel       | 1                                                     | 2 | 3 | 4 |
| Ergebnis<br>Result     | ✓                                                     | ✓ | ✓ | ✓ |

## Calibration verificate – page 3

- ▶ Reference value
- ▶ Measurement derivation
- ▶ Measurement channels / measurement ranges
- ▶ Measurement uncertainty
- ▶ **Conformity statement**

Comparison of the calibration values with limit values from specifications of the test specimen including the measurement uncertainty



# Calibration

## Factory calibration – DAkkS calibration



**CSM GmbH**  
Computer-Systeme-Messtechnik

### Kalibrierschein / Calibration certificate

erstellt durch das Kalibrierlaboratorium  
issued by the calibration laboratory

**Kalibrierlabor der CSM GmbH**

akkreditiert nach DIN EN ISO/IEC 17025:2018  
accredited according to DIN EN ISO/IEC 17025:2018

Mitglied im  
Member of

**Deutschen Kalibrierdienst DKD**



Kalibrierzeichen  
Calibration mark

127  
D-K-  
15214-01-00  
2022-06

**Gegenstand**  
Object  
**Mehrkanalmessgerät für Thermospannung**  
Typ K  
Multi-channel measurement device for thermoelectric voltage type K

**Hersteller**  
Manufacturer  
**CSM Computer-Systeme-Messtechnik GmbH**

**Typ**  
Type  
**HV TH4 evo**

**Seriennummer**  
Serial number  
**3385-HVTH4**

**Auftraggeber**  
Customer  
**CSM Computer-Systeme-Messtechnik GmbH**  
**70794 Filderstadt**  
**Germany**

**Auftragsnummer**  
Order number  
**082220084-3385**

**Anzahl der Seiten**  
Number of pages  
**- 4 -**

**Datum der Kalibrierung**  
Date of calibration  
**2022-06-27**

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der ausdrücklichen Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine sind bei Nennung des für die Freigabe Verantwortlichen in Klarschrift auch ohne Unterschrift gültig.  
This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates with the full name of the approval responsible person are valid without signature.

**Datum und Freigabe des Kalibrierscheins**  
Date and approval of the calibration certificate

signiert  
von: 27.06.2022  
um: 09:37:59 +02

Bearbeiter  
Person in charge

signiert  
von: 27.06.2022  
um: 07:49:09 +02

CSM Computer-Systeme-Messtechnik GmbH  
Kaffeehausstraße 36 · 70794 Filderstadt · Deutschland  
Es gelten ausschließlich unsere allgemeinen Geschäftsbedingungen. Der Erfüllungsort ist Filderstadt, der Gerichtsstand ist Stuttgart.  
(Only our general terms of business apply. The place of performance is Filderstadt, the place of jurisdiction is Stuttgart.)

Handelsregister: HRB 222607  
Amtsgericht Stuttgart

Seite 2/4  
Page  
D-K-  
15214-01-00  
2022-06

ent Typ K.  
DKD-R 5-5 (Abb. 3) durch  
Vergleichsstelle. Die angege-  
beziehen sich auf die Inter-  
auf 10 Hz Senderate und

DKD-R 5-5 (Fig. 3) by sensor  
calibrated from the thermoelectric  
for calibration, the test object

unter klimatisierten Bedin-  
gung.

**Kalibrierdatum**  
Calibration date  
2021-10

Multiplikation mit dem  
Größe liegt mit einer

HRB 222607  
Amtsgericht Stuttgart

Seite 3/4  
Page  
D-K-  
15214-01-00  
2022-06

stellerspezifikation des Kalibriergegenstands unter  
ungungswahrscheinlichkeit 95 %) in Anlehnung an  
undfaktor  $g = 2$ , wodurch eine Konformitätswahr-  
es wird für jeden einzelnen Messwert im Kalibrier-

Manufacturer's specification of the calibration object under  
probability 95 %) referring to DIN EN ISO 14253-1:2018-07.  
It is a conformance probability of 95 % and higher. The  
factor is as follows:

**Zeichnung**  
Drawing

**Exemplarisches Messergebnis**  
Measurement result example

| Bezeichnung | Exemplarisches Messergebnis |
|-------------|-----------------------------|
| 1           | 0,65 K                      |
| 2           | 0,80 K                      |
| 3           | 0,65 K                      |
| 4           | 1,10 K                      |
| 5           | 1,20 K                      |
| 6           | 0,65 K                      |
| 7           | 0,80 K                      |
| 8           | 0,65 K                      |
| 9           | 1,10 K                      |
| 10          | 1,20 K                      |
| 11          | 0,65 K                      |
| 12          | 0,80 K                      |
| 13          | 0,65 K                      |
| 14          | 1,10 K                      |
| 15          | 1,20 K                      |
| 16          | 0,65 K                      |
| 17          | 0,80 K                      |
| 18          | 0,65 K                      |
| 19          | 1,10 K                      |
| 20          | 1,20 K                      |
| 21          | 0,65 K                      |
| 22          | 0,80 K                      |
| 23          | 0,65 K                      |
| 24          | 1,10 K                      |
| 25          | 1,20 K                      |
| 26          | 0,65 K                      |
| 27          | 0,80 K                      |
| 28          | 0,65 K                      |
| 29          | 1,10 K                      |
| 30          | 1,20 K                      |
| 31          | 0,65 K                      |
| 32          | 0,80 K                      |
| 33          | 0,65 K                      |
| 34          | 1,10 K                      |
| 35          | 1,20 K                      |
| 36          | 0,65 K                      |
| 37          | 0,80 K                      |
| 38          | 0,65 K                      |
| 39          | 1,10 K                      |
| 40          | 1,20 K                      |
| 41          | 0,65 K                      |
| 42          | 0,80 K                      |
| 43          | 0,65 K                      |
| 44          | 1,10 K                      |
| 45          | 1,20 K                      |
| 46          | 0,65 K                      |
| 47          | 0,80 K                      |
| 48          | 0,65 K                      |
| 49          | 1,10 K                      |
| 50          | 1,20 K                      |

1 Toleranzgrenzen / Tolerance limits  
2 Referenzwert / Reference value  
3 Messwert / Measured value  
4 Messunsicherheit / Measurement uncertainty

nummer dieses Kalibrierscheins sowie  
of this calibration certificate as well as the

Multiplikation mit dem  
Größe liegt mit einer

Handelsregister: HRB 222607  
Amtsgericht Stuttgart

Handelsregister: HRB 222607  
Amtsgericht Stuttgart

Handelsregister: HRB 222607  
Amtsgericht Stuttgart

Seite 4/4  
Page  
D-K-  
15214-01-00  
2022-06

| Toleranz | Abweichung / Toleranz | Ergebnis |
|----------|-----------------------|----------|
| 0,65 K   | -26,2 %               | ✓        |
| 0,80 K   | -21,7 %               | ✓        |
| 0,65 K   | -18,5 %               | ✓        |
| 1,10 K   | -4,6 %                | ✓        |
| 1,20 K   | 1,6 %                 | ✓        |
| 0,65 K   | -12,3 %               | ✓        |
| 0,80 K   | -15,0 %               | ✓        |
| 0,65 K   | -4,6 %                | ✓        |
| 1,10 K   | -2,7 %                | ✓        |
| 1,20 K   | 0,0 %                 | ✓        |
| 0,65 K   | -15,4 %               | ✓        |
| 0,80 K   | -16,7 %               | ✓        |
| 1,10 K   | -1,5 %                | ✓        |
| 1,20 K   | 0,9 %                 | ✓        |
| 0,65 K   | 3,9 %                 | ✓        |
| 0,80 K   | -24,6 %               | ✓        |
| 0,65 K   | -23,3 %               | ✓        |
| 1,10 K   | -9,2 %                | ✓        |
| 1,20 K   | -2,7 %                | ✓        |
| 0,65 K   | 1,6 %                 | ✓        |


- ▶ Reference value
- ▶ Measurement derivation
- ▶ Measurement channels / measurement ranges
- ▶ Measurement uncertainty
- ▶ Conformity statement
- ▶ **Factory calibration**

Manufacturer conducts calibration according to own specifications



# Calibration

## Factory calibration – DAkkS calibration



Calibration certificate – page 1

**CSM GmbH**  
Computer-Systeme-Messtechnik

**Kalibrierschein / Calibration certificate**

erstellt durch das Kalibrierlaboratorium  
*issued by the calibration laboratory*

**Kalibrierlabor der CSM GmbH**

akkreditiert nach DIN EN ISO/IEC 17025:2018  
*accredited according to DIN EN ISO/IEC 17025:2018*

Mitglied im  
*Member of*

**Deutschen Kalibrierdienst** **DKD**

**DAkkS**  
Deutsche  
Akkreditierungsstelle  
D-K-15214-01-00

Kalibrierzeichen  
*Calibration mark*

|                 |
|-----------------|
| 127             |
| D-K-15214-01-00 |
| 2022-06         |

**CSM Computer-Systeme-Messtechnik GmbH**  
Raffelsteinstraße 36-70794-Flörsdorf, Deutschland  
Es gelten ausschließlich unsere allgemeinen Geschäftsbedingungen. Der Erfüllungsort ist Flörsdorf, der Gerichtsstand ist Stuttgart.  
(Our general terms of supply apply. The place of performance is Flörsdorf, the place of jurisdiction is Stuttgart.)

**Bearbeiter**  
Person in charge

signiert  
von: 27.06.2022  
um: 07:49:09 +02

Handelsregister: HRB 222607  
Amtsgericht Stuttgart

Handelsregister: HRB 222607  
Amtsgericht Stuttgart

Handelsregister: HRB 222607  
Amtsgericht Stuttgart

- ▶ Reference value
- ▶ Measurement derivation
- ▶ Measurement channels / measurement ranges
- ▶ Measurement uncertainty
- ▶ Conformity statement
- ▶ Factory calibration
- ▶ **DAkkS calibration**

ISO 17025 accredited  
calibration laboratory.  
Internationally accepted




# Calibration Interval

## Date of last calibration / next calibration

### ► Configuration software CSMconfig

(last calibration)



# Calibration Interval

## Date of last calibration / next calibration

- Configuration software CSMconfig

(last calibration)

Program Settings


|                                                    |                                     |     |
|----------------------------------------------------|-------------------------------------|-----|
| >                                                  | Measurement scheme import           |     |
| Column Heading Definitions (XML)                   | MSP\MspColumns.xml                  | ... |
| Assign S/N according to current devices            | <input checked="" type="checkbox"/> |     |
| >                                                  | ECM Related                         |     |
| Force ECM dashCANc compatibility                   | <input type="checkbox"/>            |     |
| Identifier base for ECM devices                    | 0x0181                              |     |
| Enable ECM 16-bit mode                             | <input checked="" type="checkbox"/> |     |
| Identifier base for ECM 16-bit values              | 0x0160                              |     |
| >                                                  | Calibration date monitoring         |     |
| Monitoring type                                    | on check                            | ▼   |
| Calibration interval                               |                                     | 365 |
| Lead warn time                                     |                                     | 30  |
| >                                                  | Misc.Settings                       |     |
| Display log file after bus scan.                   | <input type="checkbox"/>            |     |
| Force measurement to 0/3/6.. fractional digits     | <input checked="" type="checkbox"/> |     |
| Allow module internal adjustments                  | <input checked="" type="checkbox"/> |     |
| Unit for temperature channels                      | deg Celsius                         | ▼   |
| Temperature unit auto conversion                   | <input checked="" type="checkbox"/> |     |
| Allows to adapt settings to the device limitations | <input checked="" type="checkbox"/> |     |
| Max. number of columns in report tables            |                                     | 12  |

OK  
Cancel  
Vector Ethernet  
Export to ...  
Import


## Calibration Interval

### Date of last calibration / next calibration


- ▶ Configuration software CSMconfig (last calibration)
- ▶ Label (next calibration)



# Calibration Interval




- ▶ User determines calibration interval
  - ▶ Use case
  - ▶ Own quality requirements
  - ▶ Quality requirements by customer
- ▶ Manufacturer gives recommendation
  - ▶ CSM: Every **12 months**



# CSM Maintenance Package


- ▶ Visual inspection
- ▶ Functional test
- ▶ Input calibration
- ▶ If necessary: Adjustment
- ▶ If necessary: Second calibration

Adjustment usually  
only by the  
manufacturer





Calibration



DAkkS



HV Insulation Test



Time

# HV Insulation Test

## Background

- ▶ Measurements in the HV environment (high voltages and currents) pose dangers for technology and users



# HV Insulation Test

## Background


- ▶ Measurements in the HV environment (high voltages and currents) pose dangers for technology and users



Danger  
High-voltage!



- ▶ **CSM safety concept**
  - ▶ HV-safe cables




# HV Insulation Test

## Background

- ▶ Measurements in the HV environment (high voltages and currents) pose dangers for technology and users




- ▶ **CSM safety concept**
  - ▶ HV-safe cables
  - ▶ **Protected connectors**




# HV Insulation Test Background

- Measurements in the HV environment (high voltages and currents) pose dangers for technology and users



- CSM safety concept**

- HV-safe cables
- Protected connectors
- Isolation barrier in the measuring device**




# HV Insulation Test

## Background

- ▶ Measurements in the HV environment (high voltages and currents) pose dangers for technology and users



- ▶ **CSM safety concept**
  - ▶ HV-safe cables
  - ▶ Protected connectors
  - ▶ Isolation barrier in the measuring device
- ▶ Due to aging processes, **regular testing** of the HV insulation is necessary



# HV Insulation Test Procedure



- ▶ Norm EN 61010-1:2020
- ▶ Test of flashover strength between potentials on HV measuring device
  - ▶ All measuring channels against each other
  - ▶ All measuring channels against housing (SELV)



DUT


# HV Insulation Test Procedure



- ▶ Norm EN 61010-1:2020
- ▶ Test of flashover strength between potentials on HV measuring device
  - ▶ All measuring channels against each other
  - ▶ All measuring channels against housing (SELV)

## Test specification:

- ▶ 3,100 V DC Test voltage
- ▶ 5 s Rise time
- ▶ 5 s holding time



# HV Insulation Test Test Report

- ▶ Norm EN 61010-1:2020
- ▶ Test of flashover strength between potentials on HV m
  - ▶ All measuring channels against each other
  - ▶ All measuring channels against housing (SELV)

## Test specification:

- ▶ 3,100 V DC Test voltage
- ▶ 5 s Rise time
- ▶ 5 s holding time



## Prüfbericht HV-Isolationstest Test report High Voltage Isolation Test

Berichtsnummer 5807  
Report number 2022-09

### Prüfgegenstand / Device under test

Typ HV THMM 4  
Type  
Seriennummer 2871-HVTH4  
Serial number  
Hersteller CSM Computer-Systeme-  
Messtechnik GmbH  
Manufacturer  
Datum des Tests 2022-09-07  
Date of test  
Anzahl Seiten 1  
Number of pages

Dieser Prüfbericht ist ohne Unterschrift nicht gültig. Es gilt immer die deutsche Textfassung.  
This test report is not valid without signature. In case of doubt, the German text always applies.

### Prüfverfahren / Test procedure

Die Prüfung wurde in Anlehnung an EN61010-1:2020-03, Anhang F durchgeführt.  
The test has been carried out in accordance with EN61010-1:2020-03, Annex F.

|                                 |                 |                           |     |                        |     |
|---------------------------------|-----------------|---------------------------|-----|------------------------|-----|
| Prüfspannung<br>Testing voltage | A--B: 3100 V DC | Anstiegszeit<br>Rise time | 5 s | Haltezeit<br>Hold time | 5 s |
|---------------------------------|-----------------|---------------------------|-----|------------------------|-----|

### Prüfergebnisse / Test results

Die oben angegebene Prüfspannung wurde jeweils zwischen den in folgender Tabelle mit A und B gekennzeichneten Bereichen des Prüflings angelegt. Die orangefarbige Hinterlegung markiert den Bereich bzw. Anschluss, an den das hohe elektrische Potential angelegt wurde. Der Bericht stellt den Zustand des Gegenstands zum Zeitpunkt des Tests dar.  
The test voltage specified above was applied between the areas of the device under test indicated by A and B in the following table. The orange filling marks the area or connection to which the high electrical potential was applied. The report represents the state of the object at the time of test.

| Test Nr.<br>Test No. | Ergebnis<br>Result | SELV | CH 1 | CH 2 | CH 3 | CH 4 |
|----------------------|--------------------|------|------|------|------|------|
| 1                    | ✓                  | A    | B    | B    | B    | B    |
| 2                    | ✓                  | A    | B    | A    | A    | A    |
| 3                    | ✓                  | A    | A    | B    | A    | A    |
| 4                    | ✓                  | A    | A    | A    | B    | A    |
| 5                    | ✓                  | A    | A    | A    | A    | B    |

Datum und Freigabe des Testzertifikats  
Date and approval of the test certificate

Bearbeiter  
Person in charge

# HV Insulation Test Test Report

- ▶ Norm EN 61010-1:2020
- ▶ Test of flashover strength between potentials on HV m
  - ▶ All measuring channels against each other
  - ▶ All measuring channels against housing (SELV)

## Test specification:

- ▶ 3,100 V DC Test voltage
- ▶ 5 s Rise time
- ▶ 5 s holding time



## Prüfbericht HV-Isolationstest Test report High Voltage Isolation Test

Berichtsnummer 5807  
Report number 2022-09

### Prüfgegenstand / Device under test

Typ HV THMM 4  
Type  
Seriennummer 2871-HVTH4  
Serial number  
Hersteller CSM Computer-Systeme-  
Messtechnik GmbH  
Manufacturer  
Datum des Tests 2022-09-07  
Date of test  
Anzahl Seiten 1  
Number of pages

Dieser Prüfbericht ist ohne Unterschrift nicht gültig. Es gilt immer die deutsche Textfassung.  
This test report is not valid without signature. In case of doubt, the German text always applies.

### Prüfverfahren / Test procedure

Die Prüfung wurde in Anlehnung an EN61010-1:2020-03, Anhang F durchgeführt.  
The test has been carried out in accordance with EN61010-1:2020-03, Annex F.

Prüfspannung A--B: 3100 V DC  
Testing voltage  
Anstiegszeit 5 s  
Rise time  
Haltezeit 5 s  
Hold time

### Prüfergebnisse / Test results

Die oben angegebene Prüfspannung wurde jeweils zwischen den in folgender Tabelle mit A und B gekennzeichneten Bereichen des Prüflings angelegt. Die orangefarbige Hinterlegung markiert den Bereich bzw. Anschluss, an den das hohe elektrische Potential angelegt wurde. Der Bericht stellt den Zustand des Gegenstands zum Zeitpunkt des Tests dar.  
The test voltage specified above was applied between the areas of the device under test indicated by A and B in the following table. The orange filling marks the area or connection to which the high electrical potential was applied. The report represents the state of the object at the time of test.

| Test Nr.<br>Test No. | Ergebnis<br>Result | SELV | CH 1 | CH 2 | CH 3 | CH 4 |
|----------------------|--------------------|------|------|------|------|------|
| 1                    | ✓                  | A    | B    | B    | B    | B    |
| 2                    | ✓                  | A    | B    | A    | A    | A    |
| 3                    | ✓                  | A    | A    | B    | A    | A    |
| 4                    | ✓                  | A    | A    | A    | B    | A    |
| 5                    | ✓                  | A    | A    | A    | A    | B    |

Datum und Freigabe des Testzertifikats  
Date and approval of the test certificate

Bearbeiter  
Person in charge

# HV Insulation Test Interval


- ▶ Depending on the load of the device
- ▶ Manufacturer gives recommendation
  - ▶ CSM: Every **12 months**



Insulation test in combination with calibration in CSM maintenance package




Calibration



DAkkS



HV Insulation Test




Time



# HV Insulation Test Execution



- ▶ Manufacturer of the HV measuring device
- ▶ User with **HV Isolation Testing Station**




# HV Insulation Test Execution

- ▶ Manufacturer of the HV measuring device
- ▶ User with **HV Isolation Testing Station**



PC with test and control software



HV voltage generator

CH 1  
CH 2  
CH 3  
CH 4

## HV ISO Self-Test Adapter



- ▶ Self-testing of the entire test setup (cables and connectors)

HV Isolation Testing Station  
at [www.csm.de](http://www.csm.de)



# HV Insulation Test Execution


- ▶ Manufacturer of the HV measuring device
- ▶ User with **HV Isolation Testing Station**



# HV Insulation Test

## Execution

- ▶ Manufacturer of the HV measuring device
- ▶ User with **HV Isolation Testing Station**



## Benefits with own testing

- ▶ Devices remain on site
- ▶ Shorter test intervals
- ▶ Testing of HV-safe cables
- ▶ Cost advantage with high number of measurement modules

Similar option for calibration  
by the user



Calibration



DAkkS



HV Insulation Test

## Checklist maintenance HV measurement technology

- ▶ Keeping an eye on inspection intervals
  - ▶ Label on housing
  - ▶ Readout function in CSMconfig
- ▶ Timely notification of upcoming test to manufacturer or service provider
- ▶ Allow time for shipping
- ▶ Spare modules
- ▶ Backup of measurement configuration before shipment
- ▶ Benefit from price advantages when booking a package (calibration + HV iso test)
- ▶ In case of large number of measurement modules: own HV Isolation Testing Station

## About CSM

CSM has been setting technological standards for decentralized measurement technology in vehicle development for over 35 years. Our CAN bus and EtherCAT® measurement devices support worldwide renowned vehicle manufacturers, suppliers and service providers in their developments.

Continuous innovation and long-term satisfied customers are our guarantee for success. Together with our partner Vector Informatik, we have developed an easily scalable and powerful E-Mobility Measurement System for hybrid and electric vehicles and are constantly expanding the areas of application. With our high-voltage safe measurement systems designed for fast and synchronous measurements and power analyses, we actively accompany the change to **E-Mobility**.

**CSM GmbH** (Germany, International)

Raiffeisenstraße 36

70794 Filderstadt

Phone: +49 711 - 77 96 40

email: [sales@csm.de](mailto:sales@csm.de)


**CSM Products, Inc. USA** (USA, Canada, Mexico)

1920 Opdyke Court, Suite 200

Auburn Hills, MI 48326

Phone: +1 248 836-49 95

email: [sales@csmproductsinc.com](mailto:sales@csmproductsinc.com)



For more information and the current dates  
of CSM Xplained, please visit

[www.csm.de/webseminars](http://www.csm.de/webseminars)



**CSM Xplained**  
measurement technology